

# Optimising ergonomics in minimally invasive gynaecological surgery: a comprehensive review and practice recommendations

 Dimitrios Balafoutas<sup>1,2</sup>,  Ralf Joukhadar<sup>2</sup>,  Nikos Vlahos<sup>1</sup>

<sup>1</sup>2<sup>nd</sup> Department of Obstetrics and Gynaecology, University of Athens, Aretaieion Hospital, Athens, Greece

<sup>2</sup>Department of Obstetrics and Gynaecology, Spital Männedorf, Zurich, Switzerland

## ABSTRACT

**Background:** Modern minimally invasive gynaecological surgery greatly contributes to women's health; however, it can be physically demanding for surgeons. A plethora of available data shows that the optimisation of ergonomics in the operating room (OR) is crucial for the health and efficiency of surgeons.

**Objectives:** To provide an overview of the importance of ergonomics and clinically useful, concise recommendations.

**Methods:** A literature review with critical analysis of available data.

**Main Outcome Measures:** Impact of ergonomics on the prevalence of musculoskeletal disorders (MSDs), fatigue levels, efficiency and subjective comfort among surgeons.

**Results:** Evidence suggests that MSDs are highly prevalent among minimally invasive gynaecological surgeons and that several ergonomic interventions can greatly reduce muscle strain and improve clinical practice, with the most important being the planning of brief intraoperative breaks, the selection of proper laparoscopic instruments and the positioning of the operating table and monitor at the correct height. The adoption of robotic surgery can also improve surgical ergonomics. Clinical practice recommendations for ergonomic improvement in gynaecological laparoscopy based on the existing evidence are provided.

**Conclusions:** Surgeons must be aware of the optimal ergonomic settings in the OR and impose measures to reduce risks and achieve a comfortable environment.

**What is New?** A comprehensive, praxis-oriented review with exact ergonomic advice for minimally invasive gynaecological surgeons.

**Keywords:** Ergonomics, laparoscopic, musculoskeletal disorders, operative setting, robotic, surgeon health, surgical efficiency

## Introduction

Minimally invasive gynaecological surgery is currently used for the diagnosis and treatment of various disorders. Despite its benefits for the patients, this approach can be physically demanding and can lead to musculoskeletal injuries among surgeons,

nurses, and other healthcare workers.<sup>1</sup> Therefore, interventions that reduce these risks are needed. Ergonomics is the science of designing and arranging the workplace, equipment, and tasks to fit the capabilities and limitations of the human body. In the context of laparoscopy, ergonomics can play a

**Corresponding Author:** Dimitrios Balafoutas, MD, <sup>2</sup>nd Department of Obstetrics and Gynaecology, University of Athens, Aretaieion Hospital, Athens, Greece; Department of Obstetrics and Gynaecology, Spital Männedorf, Zurich, Switzerland

**E-mail:** dbalafoutas@yahoo.com **ORCID ID:** orcid.org/0000-0002-6256-2127

**Received:** 18.10.2024 **Accepted:** 20.04.2025 **Epub:** 24.06.2025

**Cite this article as:** Balafoutas D, Joukhadar R, Vlahos N. Optimising ergonomics in minimally invasive gynaecological surgery: a comprehensive review and practice recommendations. Facts Views Vis Obgyn. [Epub Ahead of Print]



crucial role in reducing the physical strain and improving the performance of surgical teams.<sup>2</sup> This is crucial for surgeons' career longevity and quality of life, because performing laparoscopic surgery has been shown to cause fatigue, strain and injury irrespective of age, experience and handedness.<sup>3</sup>

This review aims to explore the current knowledge on ergonomics for gynaecological laparoscopy. We investigate the hypothesis that specific ergonomic interventions can reduce the prevalence of musculoskeletal disorders (MSDs) among surgeons and improve their overall surgical performance. Moreover, we summarise concise recommendations regarding the optimal ergonomic settings based on available data.

## Methods

### **Search Strategy and Study Eligibility**

The systematic search was conducted in the ScienceDirect, PubMed/Medline, and Google Scholar databases without any restriction on the publication date. The preferred reporting items for systematic reviews and meta-analyses (PRISMA) guidelines were used.<sup>4</sup> The protocol for this review was registered at PROSPERO (CRD42023452153). The search focused on studies that evaluated ergonomic or surgeon strain parameters during laparoscopic surgery. The following keywords were used: (laparoscopic OR robotic) AND (ergonomics OR ergonomic) AND (gynecological OR gynecologic), (laparoscopic OR robotic) AND musculoskeletal AND (gynecological OR gynecologic). The search was performed in October and November 2023.

Inclusion criteria were surgeons as subjects (primary operators and assistants). Exclusion criteria were work-related MSDs among hospital staff outside the operating room (OR). The main outcomes to be considered were the avoidance of musculoskeletal injury of surgeons, the reduction in fatigue and the improved efficiency and operative time. Randomised controlled trials and prospective or retrospective randomised cohort studies were included. Because of the narrative character of this review and the need to suggest optimal ergonomic recommendations and ideas, review articles and society guideline websites were also included. Only full-text articles were included. Inaccessible articles and articles in languages other than English were excluded.

### **Data Extraction**

Articles from the initial database search were searched for duplicates. Two hundred thirty-four articles were

screened by titles and abstracts for irrelevant articles. After assessing content according to the inclusion/exclusion criteria, articles were scanned by the authors for relevant information and supplemented with online scientific committee sources and two book chapters. Finally, 86 sources were included. Study design, baseline characteristics, modality of surgery (laparoscopic or robotic) and exact setting, OR table height, OR setup and surgeon positioning were extracted for comparison from full-text articles. No missing data was defined.

### **Strategy For Data Synthesis**

Narrative synthesis assessing the quality of studies and bias.

### **Evidence**

**Research Tools:** Research tools that have been utilised to study surgical ergonomics can be broadly categorised into subjective and objective instruments.<sup>5</sup> Subjective tools include validated questionnaire scales that study discomfort in specific body regions or subjective assessment of the mental and physical workload, performance and frustration.<sup>6</sup> Objective tools include electromyography measurements of muscle activity and fatigue<sup>7</sup> and kinematic tracking through video<sup>8</sup> or special sensors, like accelerometers.<sup>9</sup> With the above research tools, valuable information about the ergonomic risk factors, as well as the common musculoskeletal problems in surgeons, could be obtained.

### **Ergonomic Risk Factors in Gynaecological Laparoscopy:**

Laparoscopy requires surgeons and staff to maintain prolonged static postures, awkward body positions, and repetitive movements, which can result in MSDs such as neck pain, back pain, shoulder pain, and hand-arm vibration syndrome. A systematic review showed that the average prevalence of physical complaints among laparoscopic surgeons was 74% and that the prevalence of MSDs is higher in minimally invasive surgeons than in any other occupational group.<sup>10</sup> Several task-related factors affect the risk for MSDs in laparoscopy, such as instrument design, equipment placement, and surgical technique, as well as individual factors, such as age, gender, and physical fitness.

Gynaecological surgeons are especially prone to MSDs because of the additional musculoskeletal strain due to the parallel exposure to vaginal surgery.<sup>11</sup> During vaginal surgery, the assistant stands holding retractors beside the primary surgeon with excessive trunk rotation and prolonged asymmetrical upper extremity strain. A study

comparing the frequency and duration of strenuous body postures between assistant and primary surgeons demonstrated that while both experience high durations of trunk lateral bending and neck and shoulder deviations, the assistant surgeons spent a greater percentage of working time in trunk flexion compared to the primary surgeon.<sup>12</sup> In operative laparoscopy, data suggest that surgical assistants face significant ergonomic stress, just as operating surgeons do.<sup>13</sup>

Many instruments, common in advanced minimally invasive gynaecological surgery, i.e. endoscopic needle-drivers, demonstrate reduced degrees of freedom, enhanced fulcrum effect, and magnification of minimal tremor.<sup>14</sup> Moreover, conventional laparoscopic instruments create an inefficient transfer of force and an uneven lever effect towards the fingers of the surgeon, which can result in pain, fatigue, and neuropraxia.<sup>15</sup>

Minimally invasive surgery involves more internal shoulder rotation, elbow flexion and wrist supination than open surgery, and larger ranges of motion are required of the upper extremities due to the instrument length.<sup>16</sup> A quantitative study of laparoscopic surgeons' movements in live surgical environment utilizing video analyses demonstrated that surgeons spent a median of 98 % (range 77-100%) of surgical time with their neck rotated at  $>21^\circ$  (range  $0^\circ$ - $52^\circ$ ) with shoulder flexion at  $45^\circ$ - $90^\circ$  for 35% vs. 0% ( $P<0.001$ ) and elbow flexion at  $>120^\circ$  for 31 vs. 0 % ( $P<0.001$ ) of total surgical time.<sup>17</sup> The non-dominant arm was subjected to more extreme positions for significantly longer periods of time compared to the dominant arm. Power morcellation was associated with the additional strain of multiple instrument insertions and removals, however, this technique is used less in recent years in many parts of the world following considerations of cancer cell dissemination. Short heighted surgeons, in particular (reference height 170 cm), spend more time in these extreme joint and posture positions.<sup>18</sup>

Hand size significantly affects the ergonomics of laparoscopic instruments and can lead to an increased risk of MSDs.<sup>19</sup> Available data suggest that smaller hand dimensions and glove size, as well as female sex, are associated with a higher probability of MSDs.<sup>14</sup> Indeed, various endoscopic surgery instruments, i.e. staplers, are designed for a minimum hand size. A study furthermore reported that the most appropriate instrument size for surgeons with a given hand size is not the same for male and female individuals, but needs to be established

separately for each sex, ideally by developing smart instruments whose usability is not affected by the gender of the user.<sup>20</sup> Unfortunately evidence suggests that this also applies to the current disposable laparoscopic devices that do not fit the needs of female laparoscopic surgeons.<sup>21</sup> Indeed, women are still more likely to describe the laparoscopic instruments as uncomfortable to handle and seek more frequent treatment for MSDs. In a recent study, women were found to have 5.37 times the odds of physical complaints attributed to the use of laparoscopic instruments (odds ratio: 5.37; 95% confidence interval: 2.56-11.25).<sup>22</sup> Because of the rapidly increasing number of women entering the field of operative gynaecology, these limitations are likely to gain importance in the future.

### **Common Musculoskeletal Disorders in Surgeons**

The overall risk of work-related musculoskeletal symptoms in surgeons has been calculated at up to 90%.<sup>23,24</sup> The highest levels have been recorded among surgeons who perform complex minimally invasive gynaecological surgery,<sup>25</sup> with 52% of the individuals reporting persistent pain in an online survey. The neck, shoulders, and wrists are the most investigated areas for MSDs, followed by the ankle, knee, back, upper back, elbow, lower back, thumbs, mid-back, fingers, and hips.<sup>26</sup> Interestingly, the prevalence of MSDs seems to increase with the number of years of laparoscopic practice.<sup>27</sup>

### **Neck and Shoulder Pain**

Neck and shoulder pain are common complaints among surgeons, with studies reporting prevalence rates ranging from 56% to 85%.<sup>28</sup> The repetitive use of upper extremities during surgery, the prolonged static postures, and the awkward positioning are all risk factors for developing neck and shoulder pain. The ergonomic impact of laparoscopy on surgeons has been studied at the level of specific muscles through electromyograms. The activation patterns of deltoid, trapezius, biceps, pronator teres, flexor carpi ulnaris, and extensor digitorum superficialis muscles have been analysed during simulated laparoscopic tasks. Proximal arm and shoulder muscles were impacted the most.<sup>29</sup>

### **Low Back Pain**

Low back pain is another common musculoskeletal complaint among surgeons. A descriptive, cross-sectional study showed prevalence rates of up to 68%.<sup>28</sup> The prolonged standing or sitting in awkward positions

during surgery, as well as the repetitive nature of surgical tasks, can contribute to the development of low back pain. Currently, limited evidence shows that exercise programs can reduce the prevalence of pain, however, most surgeons experience ongoing symptoms.<sup>30</sup>

### **Carpal Tunnel Syndrome**

Carpal tunnel syndrome is a common hand and wrist injury among surgeons, with prevalence rates up to 34%. Repetitive hand movements, awkward hand positions, and forceful gripping of instruments are all risk factors for developing carpal tunnel syndrome. An online questionnaire study found that, while ergonomic interventions, such as adjustable instrument handles and padded gloves, could reduce the incidence of carpal tunnel syndrome, most surgeons were unaware of the possible ergonomic solutions and didn't consider adopting any appropriate preventive measures.<sup>31</sup>

### **Lower Extremities**

Posture-related MSDs of the lower extremities, especially in the knee and ankle/foot regions, appear to be common among surgeons, with reported prevalence up to 65%.<sup>32</sup> Increased prevalence of varicose veins has been well-documented<sup>33</sup> and standing places significant pressure on the joints of the hips, knees, ankles and feet and without significant movement, the lubrication of the synovial joints is diminished, causing increased wear. These MSDs are of particular importance for the surgeons' quality of life, because they appear to have a maximum impact on their leisure activities.<sup>1</sup>

Interestingly, the MSDs experienced by surgeons seem to have implications on clinical practice, with up to 30% of surgeons reporting that they consider their symptoms as a factor in choosing the operative approach.<sup>34</sup>

### **Ergonomic Interventions for Gynaecological Laparoscopy**

Ergonomic interventions across a diverse range of industries in modern working environments have been shown to decrease lost workdays and sick leave,<sup>35</sup> and to improve efficiency and employee satisfaction.<sup>36</sup> In general terms, ergonomic improvements in the occupational setting have been proven to be cost-efficient<sup>37</sup> Despite this evidence, limited ergonomic interventions have been implemented for surgeons until recently.<sup>38</sup>

Ergonomic interventions can help reduce the physical strain and MSDs associated with laparoscopy. Fortunately, there are available effective ergonomic guidelines which

are proven to reduce the risk of MSDs.<sup>39</sup> Some of the commonly used ergonomic interventions in laparoscopy include the following:

**Intraoperative Breaks:** During training and clinical practice, surgeons often develop a high level of concentration on patient outcomes, which frequently leads to neglecting their own needs during operations. Therefore, even microbreaks of some seconds are uncommon in laparoscopic surgery. However, current data suggest that work breaks during complex laparoscopic surgery can reduce psychological stress and preserve performance without prolongation of the operation time compared with the traditional work scheme. A randomised clinical trial found that regular intraoperative breaks did not prolong the operation (176 vs. 180 min,  $P>0.05$ ) and the surgeon's cortisol levels, as an indicator of stress during the operation, were reduced by  $22 \pm 10.3\%$  ( $P<0.05$ ).<sup>40</sup> Another prospective study concluded that muscular fatigue and loss of accuracy can almost completely be prevented by microbreaks: In an experiment with surgeons under increasing fatigue, manual accuracy, measured by mistakes made when following a predetermined path on a board and discomfort, measured by a visual analogue scale, were vastly eliminated by microbreaks.<sup>41</sup> In a multi-centre cohort study, discomfort in the shoulders of surgeons incorporating microbreaks was significantly reduced, while distractions and flow impact were minimal, with the majority of surgeons reporting that they would alter their clinical routine after the exposure to the study.<sup>42</sup>

Regarding surgeon body positioning during prolonged laparoscopy, avoiding prolonged extreme body and trunk positions seems to be crucial. Laparoscopic surgery allows for more head/neck positioning flexibility in comparison with open surgery because the monitors can be adjusted. Preferably, the neck should have a small degree of flexion from  $15^\circ$  to  $25^\circ$ , while the shoulders should be below  $20^\circ$  of abduction and  $40^\circ$  of internal rotation.<sup>43</sup> The elbows should have a flexion of  $90^\circ$ - $120^\circ$ , and the wrists should not exceed  $15^\circ$  of deviation or flexion in any direction.<sup>44</sup> The positioning of foot pedals should be placed in an ergonomically favourable position, directly to the side of the working foot and should enable the knees to be soft and unlocked, feet hip-width apart, and body weight equally distributed. Surgeons should limit foot dorsiflexion to below  $25^\circ$  over the pedal and, if possible, utilise shoes without extreme external width, which can minimise the risk of accidental pedal and energy engagement.


The Alexander technique, a process of psychophysical re-education of the body to improve postural balance and coordination initially described in open surgery, has also been adopted in operative laparoscopy with positive impact in ergonomics and, interestingly, also in laparoscopic skills assessment scales.<sup>45</sup> The optimal body positioning for gynaecological laparoscopy is shown in Figure 1.

The design of laparoscopic instruments and equipment can significantly impact the physical strain and the performance of surgeons and staff. Ergonomically designed instruments, such as those with angled handles, adjustable tension and ergonomic grips, can reduce the strain on the hand, wrist, and forearm muscles and improve the precision and control of surgical movements. Especially those that minimise wrist flexion and rotation, and ulnar deviation should be selected.<sup>46</sup> Equipment placement, such as the position of monitors, can also impact the posture and neck flexion of surgeons and staff. Additionally, the selected instruments should be appropriate for the surgeon's anthropometry and the exact intended task.<sup>47</sup> Laparoscopic suturing and knotting constitute a special ergonomic challenge, where the camera angle and the distance between the working trocars play a crucial role. The ideal geometry has been proposed in an in vitro model study. An isosceles triangle between the instruments, with an angle between 25°

and 45° and an angle of <55° between the instruments and the horizontal, facilitates faster and more relaxed suturing.<sup>48</sup>

In recent years, handheld robotic laparoscopic instruments have been developed. While lacking the motorised arm support of the full-scale robotic platforms, these instruments aim to improve ergonomics in complex laparoscopic tasks like intracorporeal suturing.<sup>49</sup> Indeed, the design of these instruments enables up to 360° rotation and some degree of three-dimensional articulation and can be combined with several end effectors, possibly reducing prolonged awkward wrist positions for the surgeon.<sup>50</sup> Furthermore, proximal interphalangeal flexion of the thumb and the metacarpophalangeal and proximal interphalangeal flexion of the index finger seem to be reduced with handheld robotic assistance.<sup>51</sup>

**Structured Training and Education:** Proper training and education can improve the ergonomic awareness and skills of surgical teams and reduce the risk of MSDs. Training programs can include instruction on proper body mechanics, postures, and movements, as well as exercises to improve strength, flexibility, and endurance. A recent electromyography study found that trained individuals had lower muscle activation ( $P<0.05$ ), muscle workload ( $P<0.05$ ) and better bimanual dexterity than the trainee surgeons at baseline.<sup>52</sup>



**Figure 1.** Recommended posture and setting for gynaecological laparoscopy.

**Environmental Modifications:** Environmental modifications, such as adjustable lighting, temperature control and noise reduction, can improve the comfort and well-being of surgical teams and reduce the risk of MSDs. Modern laparoscopic ORs are equipped with multiple ceiling-suspended flat-screen monitors that facilitate versatile positioning around the operative field. The exact adjustment of each monitor in location, height, and inclination within a comfort distance and in the direct field of vision of each surgeon can reduce eyestrain and improve posture during prolonged operations.<sup>44,53</sup> The correct placement of the endoscopic image, as a sole intervention, has been shown to decrease the operative time by 10%, even for procedures that do not require complex suturing skills.<sup>54</sup> In the case of intracorporeal knot tying, a randomised controlled trial could demonstrate that both knot quality ( $P<0.01$ ) and execution times ( $P<0.01$ ) could be improved with the monitor straight in front of the operator at the level of the hands.<sup>55</sup> This finding contrasts with the common perception of the OR staff that the monitor should be at the level of the eyes or higher. Indeed, the optimal height zone appears to lie 15 degrees lower than sight level. The direct ergonomic impact of monitor positioning could be highlighted in a study utilising electromyography and ultrasonic position transmitters, which compared several monitor angles (display at 0°, 17.5°, and -35°) and clearly proved that muscle effort increased with viewing angle.<sup>56</sup>

Recent data suggest that proper workload management, such as task rotation, can reduce the physical strain and fatigue associated with laparoscopy. Task rotation can help distribute the physical demands across different body regions and reduce the physical strain.<sup>57,58</sup> In particular, surgeons and assistants switching sides of the table to balance the strain on the upper extremities has been proposed.<sup>59</sup>

Proper holding and manipulation of laparoscopic instruments are essential for successful laparoscopic surgery. Incorrect handling of the instruments can lead to tissue damage, prolong the surgery, and increase the risk of complications. The surgeon should hold the laparoscopic instrument in a relaxed and comfortable grip, using the thumb and index finger. The grip should be firm enough to control the instrument, but not so tight as to cause hand fatigue. The other fingers should be relaxed and not holding the instrument, as this can cause unnecessary tension and strain.<sup>60</sup> Using the dominant hand can improve the surgeon's dexterity and control over

the instrument, reducing the risk of tissue damage and other complications. The surgeon should use their wrist and fingers to manipulate the laparoscopic instrument, rather than their shoulder or elbow. This can reduce the risk of shoulder and neck strain, as well as improve the surgeon's control over the instrument (fine positioning).<sup>61</sup>

The height of the operating table is an important factor to consider during laparoscopic surgery, as it can affect the surgeon's posture and increase the risk of musculoskeletal injuries. The optimal height of the operating table for laparoscopic surgery depends on several factors, including the surgeon's height, the type of procedure, and the size of the patient. Generally, the operating table height should be adjusted to ensure that the surgeon's elbows are at a comfortable and neutral position when holding laparoscopic instruments.<sup>62</sup> OR tables were designed for open operations and are too high for many surgeons performing laparoscopic surgery. The ergonomically optimal operating surface height for laparoscopic surgery has been previously assessed in a study performed in a pelvic-trainer setting, with the strain being measured with questionnaires and electromyography.<sup>44</sup> The optimal patient height during a laparoscopic procedure is suggested to be 0.7× to 0.8× surgeon elbow height, which allows joints to stay in their neutral position for more than 90% of the operation duration. This proposed formula results in heights with an average of only 77 cm, whereas for open surgery, the equivalent lies at about 122 cm. Usual operating tables have a range of 73-122 cm, which, given the extra height of the supine patient, would be too high for 95% of minimally invasive surgeons.<sup>63</sup> While a stool is available in every setting, this solution is not sufficient in all scenarios. Energy devices require the surgeon's pedals and balance of the surgeon, and with the parallel use of various pedals, can be demanding.

**Special Equipment which Aims Solely to Improve Surgeons' Comfort is Available:** Special ergonomic chairs with adjustable heights should be readily available.<sup>64</sup> For prolonged operations, a randomised controlled trial has shown that robot-assisted camera holders can decrease the strain of the assistants.<sup>65</sup> The OR staff should ensure that the lights are adequately dimmed to ensure glare reduction and display contrast enhancement, while simultaneously allowing safe movements throughout the room.<sup>66</sup> Cables and tubes usually clutter the floor of the OR, creating physical hazards for operators and staff. Organising the cables at the beginning of surgery, as well

as ceiling-mounted boom systems for cables outside of the direct proximity of surgeons, can enhance safety and reduce physical obstacles, hence improving ergonomics. Whereas it has been proven that surgeons can effectively block out noise, it is preferable to reduce noise in the OR to improve communication within the team, especially in emergencies.<sup>67</sup> Additionally, when planning ergonomics for complex gynaecological laparoscopy, it is important to organise both patient and equipment placement to facilitate conversion to laparotomy or patient resuscitation.

### **Ergonomic Factors of Robotic Surgery**

Robotic surgery is a minimally invasive surgical technique that uses robotic systems to perform surgical procedures. It offers several ergonomic benefits over traditional open or laparoscopic surgery, which can improve surgical outcomes and reduce the risk of injuries for the surgical team. At the same time, robotic surgery creates new challenges and special issues that must be addressed.

The customizability of the surgeon's console can greatly improve surgeon ergonomics, resulting in less overall back, shoulder, neck, and wrist pain.<sup>68</sup> A recent prospective cohort study suggested adjusting the console to achieve the most neutral neck angle and lowering the viewfinder until visibility into the device is uninhibited while sitting up straight, usually at a viewing angle of approximately 15° below the horizontal.<sup>69</sup> Back flexion should be less than 15°, while neck flexion should not exceed 25°, which is a low-risk posture as assessed in MSDs risk assessment validated tools.<sup>8</sup> Robotic surgeons should be instructed that the head should rest lightly on the console headrest to avoid forehead pain and increased neck strain.<sup>70</sup> Forearms should rest on the console armrests to cater for a more relaxed soldier position and free flexion of the elbows.<sup>71</sup> It is important to frequently utilise the clutches that enable the free adjustment of the controls to keep the hands in the neutral position ("sweet spot" in the robotic surgery argot).<sup>46</sup> The recommended surgeon positioning for ergonomic improvement in robotic surgery is shown in Figure 2.

### **Reduced Physical Strain and Fatigue**

Robotic surgery systems allow for more ergonomic positioning for the surgical team, which can reduce physical strain and fatigue. The surgeon sits at a console that is typically located away from the patient, allowing for a more comfortable, neutral posture. This can reduce the risk of musculoskeletal injuries, such as neck and

back pain, which are common in traditional laparoscopic surgery. A survey of physical discomfort and symptoms following open, laparoscopic, and robotic surgery found that surgeons experienced significantly less physical strain and fatigue during robotic-assisted surgery compared to laparoscopic surgery.<sup>72</sup> Additionally, the forearms can rest on the armrest of the console and are hereby protected from gravity strain.<sup>73</sup>

### **Improved Visualization**

Robotic surgery systems offer improved visualisation of the operative field, which can reduce the risk of errors and complications. The systems provide high-definition 3D imaging, which allows for better depth perception and visualisation of anatomical structures. This can reduce the need for awkward head positions or repeated instrument exchanges and can improve ergonomics for the surgical team. Several studies found that through the tremor-free 3D immersive optics, robotic surgery provided better visualisation of the surgical field compared to laparoscopic surgery.<sup>74,75</sup>

### **More Precise Instrument Control**

Robotic surgery systems offer more precise instrument control, which can reduce the risk of errors and complications. Robotic instruments are designed to mimic the movements of the surgeon's hand and wrist, allowing for greater dexterity and control.<sup>76</sup> This can reduce the need for excessive force or repetitive motions, which can reduce the risk of injuries caused by hand and wrist strain. Currently, the use and demand for robotic medical and surgical platforms are increasing, and new technologies are continuously being developed with promising possible ergonomic advantages for surgeons.<sup>77</sup>

Importantly, MSDs persist in robotic surgery, albeit at a lower rate than in laparoscopic surgery.<sup>78</sup> In the field of gynaecology, a large survey reported 54% of participating gynaecologic robotic surgeons experiencing physical symptoms or discomfort.<sup>79</sup> Discomfort in the fingers and neck was the most reported problem. In an online questionnaire survey robotic surgery was found to be more likely than either open or laparoscopic surgery to lead to eye or finger symptoms, and more likely than open surgery (but not laparoscopic surgery) to lead to thumb symptoms.<sup>72</sup> Additionally, prolonged sitting without lumbar support creates greater intradiscal strain than standing.<sup>80</sup> A further ergonomic limitation of robotic surgeons affects bedside assistant surgeons, who are exposed to unnatural positions under the threat of

sudden motion of the robotic arms. In one study, 73% of bedside assistants reported discomfort, stressful positioning of the upper extremities, trunk, neck, and shoulder.<sup>81</sup> A further study reported that robotic assistance is associated with worse neck posture, but lower overall and mental workload compared to the console surgeon.<sup>82</sup> Importantly, a questionnaire survey reported that only a small percentage of robotic surgeons (17%) received ergonomic training prior to practice.<sup>38</sup>


In conclusion, robotic surgery offers several ergonomic benefits over traditional open or laparoscopic surgery. It allows for more ergonomic positioning for the surgical team, improved visualisation of the operative field, more precise instrument control, and reduced smoke and noise exposure. These benefits can improve surgical outcomes and reduce the risk of injuries for the surgical team. However, more studies are needed to explore the long-term effects of robotic surgery on the ergonomics and health of the surgical team.

Based on the mentioned evidence, we propose an ergonomics checklist for the minimally invasive gynaecological surgeon (Table 1) to safeguard his/her own well-being and the well-being of the surgical team.

## Discussion

Even though, when confronted with questionnaires, surgeons answer that ergonomics should be part of minimally invasive gynaecological surgery training, less than 20% of surgeons report ergonomic training during residency and fellowship, and less than two-thirds of surgeons with one-time training in ergonomics incorporate those principles into practice.<sup>83,84</sup>

Work-related MSDs have an enormous impact on work absenteeism and decreased productivity.<sup>85</sup> Moreover, they have a negative impact on the healthcare professionals quality of life.<sup>86</sup> Entering the OR, gynaecological minimally invasive surgeons follow guidelines and standard operating procedures to ensure patient safety. Unfortunately, surgeon safety has received little attention in the demanding and developing field of minimally invasive surgery, creating an environment in which "patients benefit while surgeons suffer".<sup>84</sup> Hence, we propose that proper ergonomics are integrated in the preoperative team-time-out checklists of minimally invasive gynaecological surgery. Additionally, and in this context, "we should stand by our surgical assistants"<sup>87</sup> and ensure that all our colleagues, including, in particular, the second assistant, frequently seated between the legs do have proper ergonomic conditions and unhindered vision of the monitors. In robotic surgery, care should



**Figure 2.** Recommended posture for gynaecological robotic surgery.

**Table 1.** Proposed ergonomic checklist for minimally invasive gynaecological surgery.

Patient positioning, room settings
Patient positioning, i.e. arms should not interfere with the surgeon
The operating table must be adjusted to optimise the surgeon's posture, and avoid using stools
The monitor should be slightly below eye level, at the level of hands, to maintain a neutral neck posture
<b>Instrumentation</b>
Ergonomically designed and familiar instruments, which use trigger locks and ratchets, should be used to minimise sustained gripping
<b>Surgeon positioning</b>
Keep your back straight, shoulders relaxed, and feet flat on the floor
The wrists should be straight and not bent, with the hands and fingers relaxed. when available, use the instruments' rotation
<b>Organizational</b>
Surgeons should take regular, preferably preplanned breaks during long procedures to rest and stretch their muscles
If possible, switch to robotics for complex operations
In robotics, follow the exact console instructions for ergonomic adjustment
Communicate ergonomic difficulties, encourage assistants to speak out

be taken that the assistants are not threatened by the sudden movements of the robotic arms.

The American College of Surgeons Division of Education established a Surgical Ergonomics Committee to systematically address the ergonomic challenges experienced by surgeons and improve their ergonomic well-being.<sup>66</sup> A well-documented recommendations bulletin with detailed general and technique-specific recommendations has been issued in 2022.<sup>88</sup> Worldwide, many hands-on laparoscopic training courses focus on ergonomic improvements and teach the proper OR settings.

Rehearsal of surgical techniques through simulation training enables tutors to demonstrate the appropriate posture and surgical technique as well as the correct utilisation of surgical instruments, hence significantly contributing to ergonomic improvements.<sup>89</sup> In this context, it is possible to assess ergonomics from video recordings during simulation training using automated movement assessment tools. The results can enable trainees to improve their posture and skills at the very early stages of their surgical career.<sup>90</sup>

Switching to robotic-assisted laparoscopy can be seen as an ergonomic upgrade in most scenarios. Additionally, current robotic surgical systems facilitate the central collection of real-time surgical data. These data can be analysed and, given the ability to integrate multiple sources simultaneously and the advances in artificial intelligence, console ergonomics are likely to be further

improved to fit most surgeons.<sup>91</sup> However, the availability of this infrastructure is still scarce due to cost.

This report focuses on the importance of improved ergonomics for surgeons' well-being. However, it has been shown that many incidents which affect patient safety can be attributed to poor ergonomics of healthcare personnel.<sup>92</sup> Even though there is high-quality data that demonstrates that workplace ergonomics improve outcomes, especially in healthcare, the direct effect of improvements in laparoscopy ergonomics on complication rates is yet to be measured.

In modern healthcare, financial cost arises as an important factor in decisions and planning. Providing the training, settings and infrastructure for optimal ergonomics in the high-tech setting of modern ORs will, inevitably, commit financial resources. Therefore, the decision makers acceptance of ergonomic improvements in minimally invasive gynaecological surgery will increase if this improvement proves to be cost-effective. Indeed, ergonomic interventions have proven themselves cost-effective through predictive cost-benefit analyses in most industries and can be seen as a safety intervention.<sup>93</sup> Hopefully, future regulatory changes in occupational safety will facilitate these improvements internationally.

### **Strengths and Strengths and Limitations of the Study**

There are obvious limitations in the applicability of recommendations on optimal ergonomics in minimally invasive gynaecological surgery: Exceptions should be

made to fit the anthropometric differences between surgeons or special situations such as pregnancy or obesity, as well as the target anatomy of the patient.<sup>94</sup> Additionally, some interventions will not be possible in some institutions due to financial reasons.

## Conclusion

This review has demonstrated the importance of ergonomics in minimally invasive gynaecological surgery and that general recommendations regarding ergonomic interventions are possible. Along with our commitment to the well-being of the patients, it is our responsibility as physicians to ensure optimal conditions for our working environment.

**Acknowledgments:** None.

**Contributors:** Concept: N.V., Design: D.B., N.V., Data Collection or Processing: D.B., Analysis or Interpretation: D.B., R.J., Literature Search: D.B., Writing: D.B., R.J., N.V.

**Funding:** The authors declared that this study received no financial support.

**Competing interests:** No conflict of interest was declared by the authors.

**Ethical approval:** Not required.

**Data sharing:** All data used for this study are published and publicly available.

**Transparency:** The authors affirm that the manuscript is an honest, accurate, and transparent account of the study being reported; that no important aspects of the study have been omitted; and that there were no discrepancies from the study as planned and registered.

## References

1. Rață AL, Barac S, Garleanu LL, Onofrei RR. Work-related musculoskeletal complaints in surgeons. *Healthcare (Basel)*. 2021;9:1482.
2. Lin E, Young R, Shields J, Smith K, Chao L. Growing pains: strategies for improving ergonomics in minimally invasive gynecologic surgery. *Curr Opin Obstet Gynecol*. 2023;35:361-7.
3. Cuschieri A. Whither minimal access surgery: tribulations and expectations. *Am J Surg*. 1995;169:9-19.
4. Schulz KF, Altman DG, Moher D; CONSORT Group. CONSORT 2010 statement: updated guidelines for reporting parallel group randomised trials. *BMJ*. 2010;340:c332.
5. Catanzarite T, Tan-Kim J, Whitcomb EL, Menefee S. Ergonomics in surgery: a review. *Female Pelvic Med Reconstr Surg*. 2018;24:1-12.
6. Olendorf MR, Drury CG. Postural discomfort and perceived exertion in standardized box-holding postures. *Ergonomics*. 2001;44:1341-67.
7. Koneczny S, Matern U. Instruments for the evaluation of ergonomics in surgery. *Minim Invasive Ther Allied Technol*. 2004;13:167-77.
8. McAtamney L, Nigel Corlett E. RULA: a survey method for the investigation of work-related upper limb disorders. *Appl Ergon*. 1993;24:91-9.
9. Collins SA, O'Sullivan DM, Tulikangas PK. Surgeon activity in robotic versus abdominal gynecologic surgery. *J Robot Surg*. 2012;6:333-6.
10. Gutierrez-Diez MC, Benito-Gonzalez MA, Sancibrian R, Gandarillas-Gonzalez MA, Redondo-Figuero C, Manuel-Palazuelos JC. A study of the prevalence of musculoskeletal disorders in surgeons performing minimally invasive surgery. *Int J Occup Saf Ergon*. 2018;24:111-7.
11. Kim-Fine S, Woolley SM, Weaver AL, Killian JM, Gebhart JB. Work-related musculoskeletal disorders among vaginal surgeons. *Int Urogynecol J*. 2013;24:1191-200.
12. Yurteri-Kaplan LA, Zhu X, Iglesia CB, Gutman RE, Sokol AI, Paquet V, Park AJ. Differences in postural loading between primary and assistant surgeons during vaginal surgery. *Int J Ind Ergon*. 2018;65:60-7.
13. Zihni AM, Cavallo JA, Ray S, Ohu I, Cho S, Awad MM. Ergonomic analysis of primary and assistant surgical roles. *J Surg Res*. 2016;203:301-5.
14. Shepherd JM, Harilingam MR, Hamade A. Ergonomics in laparoscopic surgery--a survey of symptoms and contributing factors. *Surg Laparosc Endosc Percutan Tech*. 2016;26:72-7.
15. Berguer R. Surgical technology and the ergonomics of laparoscopic instruments. *Surg Endosc*. 1998;12:458-62.
16. Nguyen NT, Ho HS, Smith WD, Philipps C, Lewis C, De Vera RM, et al. An ergonomic evaluation of surgeons' axial skeletal and upper extremity movements during laparoscopic and open surgery. *Am J Surg*. 2001;182:720-4.
17. Aitchison LP, Cui CK, Arnold A, Nesbitt-Hawes E, Abbott J. The ergonomics of laparoscopic surgery: a quantitative study of the time and motion of laparoscopic surgeons in live surgical environments. *Surg Endosc*. 2016;30:5068-76.
18. Manasayakorn S, Cuschieri A, Hanna GB. Ergonomic assessment of optimum operating table height for hand-assisted laparoscopic surgery. *Surg Endosc*. 2009;23:783-9.
19. Berguer R, Hreljac A. The relationship between hand size and difficulty using surgical instruments: a survey of 726 laparoscopic surgeons. *Surg Endosc*. 2004;18:508-12.
20. Kono E, Tomizawa Y, Matsuo T, Nomura S. Rating and issues of mechanical anastomotic staplers in surgical practice: a survey of 241 Japanese gastroenterological surgeons. *Surg Today*. 2012;42:962-72.
21. Adams DM, Fenton SJ, Schirmer BD, Mahvi DM, Horvath K, Nichol P. One size does not fit all: current disposable laparoscopic devices do not fit the needs of female laparoscopic surgeons. *Surg Endosc*. 2008;22:2310-3.
22. Wong JMK, Moore KJ, Carey ET. Investigation of the association between surgeon sex and laparoscopic device ergonomic strain in gynecologic surgery. *J Minim Invasive Gynecol*. 2022;29:984-91.
23. Liang CA, Levine VJ, Dusza SW, Hale EK, Nehal KS. Musculoskeletal disorders and ergonomics in dermatologic surgery: a survey of Mohs surgeons in 2010. *Dermatol Surg*. 2012;38:240-8.
24. Szeto GP, Ho P, Ting AC, Poon JT, Cheng SW, Tsang RC. Work-related musculoskeletal symptoms in surgeons. *J Occup Rehabil*. 2009;19:175-84.
25. Franasiak J, Ko EM, Kidd J, Secord AA, Bell M, Boggess JF, et al. Physical strain and urgent need for ergonomic training among gynecologic oncologists who perform minimally invasive surgery. *Gynecol Oncol*. 2012;126:437-42.

26. Gorce P, Jacquier-Bret J. Effect of assisted surgery on work-related musculoskeletal disorder prevalence by body area among surgeons: systematic review and meta-analysis. *Int J Environ Res Public Health*. 2023;20:6419.
27. Dixon F, Vitish-Sharma P, Khanna A, Keeler BD. Work-related musculoskeletal pain and discomfort in laparoscopic surgeons: an international multispecialty survey. *Ann R Coll Surg Engl*. 2023;105:734-8.
28. Alshareef L, Al Luhaybi F, Alsamli RS, Alsulami A, Alfahmi G, Mohamedelhussein WA, et al. Prevalence of back and neck pain among surgeons regardless of their specialties in Saudi Arabia. *Cureus*. 2023;15:e49421.
29. Quick NE, Gillette JC, Shapiro R, Adrales GL, Gerlach D, Park AE. The effect of using laparoscopic instruments on muscle activation patterns during minimally invasive surgical training procedures. *Surg Endosc*. 2003;17:462-5.
30. Al Amer HS. Low back pain prevalence and risk factors among health workers in Saudi Arabia: a systematic review and meta-analysis. *J Occup Health*. 2020;62:e12155.
31. Michael S, Mintz Y, Brodie R, Assalia A. Minimally invasive surgery and the risk of work-related musculoskeletal disorders: results of a survey among Israeli surgeons and review of the literature. *Work*. 2022;71:779-85.
32. Dianat I, Bazazan A, Souraki Azad MA, Salimi SS. Work-related physical, psychosocial and individual factors associated with musculoskeletal symptoms among surgeons: Implications for ergonomic interventions. *Appl Ergon*. 2018;67:115-24.
33. Qari TA, Almatrafi KN, Khateb FR, Al-Kaabi B, Al-Harbi A, Alabdali S, et al. Prevalence of varicose veins among surgeons: a cross-sectional study. *Cureus*. 2024;16:e67687.
34. Stucky CH, Cromwell KD, Voss RK, Chiang YJ, Woodman K, Lee JE, et al. Surgeon symptoms, strain, and selections: systematic review and meta-analysis of surgical ergonomics. *Ann Med Surg (Lond)*. 2018;27:1-8.
35. Rivilis I, Van Eerd D, Cullen K, Cole DC, Irvin E, Tyson J, et al. Effectiveness of participatory ergonomic interventions on health outcomes: a systematic review. *Appl Ergon*. 2008;39:342-58.
36. Hendrick HW. Determining the cost-benefits of ergonomics projects and factors that lead to their success. *Appl Ergon*. 2003;34:419-27.
37. Beevis D. Ergonomics--costs and benefits revisited. *Appl Ergon*. 2003;34:491-6.
38. Fransiak J, Craven R, Mosaly P, Gehrig PA. Feasibility and acceptance of a robotic surgery ergonomic training program. *JLS*. 2014;18:e2014.00166.
39. Alaqeel M, Tanzer M. Improving ergonomics in the operating room for orthopaedic surgeons in order to reduce work-related musculoskeletal injuries. *Ann Med Surg (Lond)*. 2020;56:133-8.
40. Engelmann C, Schneider M, Kirschbaum C, Grote G, Dingemann J, Schoof S, et al. Effects of intraoperative breaks on mental and somatic operator fatigue: a randomized clinical trial. *Surg Endosc*. 2011;25:1245-50.
41. Dorion D, Darveau S. Do micropauses prevent surgeon's fatigue and loss of accuracy associated with prolonged surgery? An experimental prospective study. *Ann Surg*. 2013;257:256-9.
42. Hallbeck MS, Lowndes BR, Bingener J, Abdelrahman AM, Yu D, Bartley A, et al. The impact of intraoperative microbreaks with exercises on surgeons: a multi-center cohort study. *Appl Ergon*. 2017;60:334-41.
43. Matern U, Waller P. Instruments for minimally invasive surgery: principles of ergonomic handles. *Surg Endosc*. 1999;13:174-82.
44. van Veelen MA, Kazemier G, Koopman J, Goossens RH, Meijer DW. Assessment of the ergonomically optimal operating surface height for laparoscopic surgery. *J Laparoendosc Adv Surg Tech A*. 2002;12:47-52.
45. Reddy PP, Reddy TP, Roig-Francoli J, Cone L, Sivan B, DeFoor WR, et al. The impact of the alexander technique on improving posture and surgical ergonomics during minimally invasive surgery: pilot study. *J Urol*. 2011;186(4 Suppl):1658-62.
46. Gabrielson AT, Clifton MM, Pavlovich CP, Biles MJ, Huang M, Agnew J, et al. Surgical ergonomics for urologists: a practical guide. *Nat Rev Urol*. 2021;18:160-9.
47. Albayrak A. Ergonomics in the operating room: transition from open to image-based surgery. 2008.
48. Frede T, Stock C, Renner C, Budair Z, Abdel-Salam Y, Rassweiler J. Geometry of laparoscopic suturing and knotting techniques. *J Endourol*. 1999;13:191-8.
49. Feng J, Yan Z, Li M, Zhang Z, Chen X, Du Z, Yang K. Handheld robotic needle holder training: slower but better. *Surg Endosc*. 2021;35:1667-74.
50. Shakir T, Chand M. Clinical applications of handheld robotic devices in general surgery: a mini-review. *Surgery*. 2024;176:1297-301.
51. Sánchez-Margallo FM, Sánchez-Margallo JA. Assessment of postural ergonomics and surgical performance in laparoendoscopic single-site surgery using a handheld robotic device. *Surg Innov*. 2018;25:208-17.
52. Khan WF, Krishna A, Roy A, Prakash O, Jaryal AK, Deepak KK, et al. Effect of structured training in improving the ergonomic stress in laparoscopic surgery among general surgery residents. *Surg Endosc*. 2021;35:4825-33.
53. Berci G, Phillips EH, Fujita F. The operating room of the future: what, when and why? *Surg Endosc*. 2004;18:1-5.
54. Erfanian K, Luks FI, Kurkchubasche AG, Wesselhoeft CW Jr, Tracy TF Jr. In-line image projection accelerates task performance in laparoscopic appendectomy. *J Pediatr Surg*. 2003;38:1059-62.
55. Hanna GB, Shimi SM, Cuschieri A. Task performance in endoscopic surgery is influenced by location of the image display. *Ann Surg*. 1998;227:481-4.
56. Bauer W, Wittig T. Influence of screen and copy holder positions on head posture, muscle activity and user judgement. *Appl Ergon*. 1998;29:185-92.
57. Choi SH, Kuchta K, Rojas A, Mehdi SA, Ramirez Barriga M, Hays S, et al. Residents perform better technically, have less stress and workload, and prefer robotic to laparoscopic technique during inanimate simulation. *Surg Endosc*. 2023;37:7230-7.
58. Krämer B, Neis F, Reisenauer C, Walter C, Brucker S, Wallwiener D, et al. Save our surgeons (SOS) - an explorative comparison of surgeons' muscular and cardiovascular demands, posture, perceived workload and discomfort during robotic vs. laparoscopic surgery. *Arch Gynecol Obstet*. 2023;307:849-62.
59. Berguer R. Ergonomics in laparoscopic surgery. In Whelan RL, Fleshman JW, Fowler DL, (eds). *The SAGES manual: perioperative care in minimally invasive surgery*. New York, NY: Springer New York. 2006;454-64.
60. Lin CJ, Chen HJ. The investigation of laparoscopic instrument movement control and learning effect. *Biomed Res Int*. 2013;2013:349825.

61. Akamatsu M, MacKenzie IS. Changes in applied force to a touchpad during pointing tasks. *Int J Indl Ergon.* 2002;29:171-82.
62. Berquer R, Smith WD, Davis S. An ergonomic study of the optimum operating table height for laparoscopic surgery. *Surg Endosc.* 2002;16:416-21.
63. Matern U, Waller P, Giebmeyer C, Rückauer KD, Farthmann EH. Ergonomics: requirements for adjusting the height of laparoscopic operating tables. *JLS.* 2001;5:7-12.
64. Schurr MO, Buess GF, Wieth F, Saile HJ, Botsch M. Ergonomic surgeon's chair for use during minimally invasive surgery. *Surg Laparosc Endosc Percutan Tech.* 1999;9:244-7.
65. Wijsman PJM, Molenaar L, Van't Hullenaar CDP, van Vugt BST, Bleeker WA, Draaisma WA, et al. Ergonomics in handheld and robot-assisted camera control: a randomized controlled trial. *Surg Endosc.* 2019;33:3919-925.
66. American College of Surgeons. (2023). Surgical Ergonomics Committee. Retrieved from <https://www.facs.org/about-acsgovernance/acs-committees/surgical-ergonomics-committee/>
67. Moorthy K, Munz Y, Undre S, Darzi A. Objective evaluation of the effect of noise on the performance of a complex laparoscopic task. *Surgery.* 2004;136:25-30; discussion 31.
68. Tarr ME, Brancato SJ, Cunkelman JA, Polcari A, Nutter B, Kenton K. Comparison of postural ergonomics between laparoscopic and robotic sacrocolpopexy: a pilot study. *J Minim Invasive Gynecol.* 2015;22:234-8.
69. Hokenstad ED, Hallbeck MS, Lowndes BR, Morrow MM, Weaver AL, McGree M, et al. Ergonomic robotic console configuration in gynecologic surgery: an interventional study. *J Minim Invasive Gynecol.* 2021;28:850-9.
70. Craven R, Franasiak J, Mosaly P, Gehrig PA. Ergonomic deficits in robotic gynecologic oncology surgery: a need for intervention. *J Minim Invasive Gynecol.* 2013;20:648-55.
71. Ronstrom C, Hallbeck S, Lowndes B, Thiels C, Bingener J. Surgical ergonomics. In: Köhler TS, Schwartz B, editors. *Surgeons as educators: a guide for academic development and teaching excellence.* Cham: Springer International Publishing; 2018. p. 387-417.
72. Plerhoples TA, Hernandez-Boussard T, Wren SM. The aching surgeon: a survey of physical discomfort and symptoms following open, laparoscopic, and robotic surgery. *J Robot Surg.* 2012;6:65-72.
73. Lux MM, Marshall M, Erturk E, Joseph JV. Ergonomic evaluation and guidelines for use of the daVinci Robot system. *J Endourol.* 2010;24:371-5.
74. Renda A, Vallancien G. Principles and advantages of robotics in urologic surgery. *Curr Urol Rep.* 2003;4:114-8.
75. Stylopoulos N, Rattner D. Robotics and ergonomics. *Surg Clin North Am.* 2003;83:1321-37.
76. Guyton SW. Robotic surgery: the computer-enhanced control of surgical instruments. *Otolaryngol Clin North Am.* 2002;35:1303-16, viii.
77. Peters BS, Armijo PR, Krause C, Choudhury SA, Oleynikov D. Review of emerging surgical robotic technology. *Surg Endosc.* 2018;32:1636-55.
78. Giberti C, Gallo F, Francini L, Signori A, Testa M. Musculoskeletal disorders among robotic surgeons: a questionnaire analysis. *Arch Ital Urol Androl.* 2014;86:95-8.
79. Lee SR, Shim S, Yu T, Jeong K, Chung HW. Sources of pain in laparoendoscopic gynecological surgeons: an analysis of ergonomic factors and proposal of an aid to improve comfort. *PLoS One.* 2017;12:e0184400.
80. Pope MH, Goh KL, Magnusson ML. Spine ergonomics. *Annu Rev Biomed Eng.* 2002;4:49-68.
81. Van't Hullenaar CDP, Bos P, Broeders IAMJ. Ergonomic assessment of the first assistant during robot-assisted surgery. *J Robot Surg.* 2019;13:283-8.
82. Yu D, Dural C, Morrow MM, Yang L, Collins JW, Hallbeck S, et al. Intraoperative workload in robotic surgery assessed by wearable motion tracking sensors and questionnaires. *Surg Endosc.* 2017;31:877-86.
83. Jensen MJ, Pagedar NA, Sugg SL, Lal G. Endocrine surgeons have high rates of work-related musculoskeletal (MSK) injury and symptoms. *Am J Surg.* 2022;224(1 Pt B):315-8.
84. Park A, Lee G, Seagull FJ, Meenaghan N, Dexter D. Patients benefit while surgeons suffer: an impending epidemic. *J Am Coll Surg.* 2010;210:306-13.
85. Bernard BP, Putz-Anderson V. Musculoskeletal disorders and workplace factors : a critical review of epidemiologic evidence for work-related musculoskeletal disorders of the neck, upper extremity, and low back. 1997.
86. Karwowski W, Marras WS. Occupational ergonomics: principles of work design: CRC Press. 2003.
87. Sakata S, Lonne MLR, Pappas CP, Stevenson ARL. We should stand by our surgical assistants. *Tech Coloproctol.* 2022;26:765-6.
88. American College of Surgeons. (2022). *Surgical Ergonomics Recommendations Division of Education & Surgical Ergonomics Committee* Retrieved from <https://www.facs.org/for-medical-professionals/education/programs/surgical-ergonomics/recommendations/>
89. Papaspyros SC, Kar A, O'Regan D. Surgical ergonomics. Analysis of technical skills, simulation models and assessment methods. *Int J Surg.* 2015;18:83-7.
90. Kratzke IM, Zhou G, Mosaly P, Farrell TM, Crouner J, Yu D. Evaluating the ergonomics of surgical residents during laparoscopic simulation: a novel computerized approach. *Am Surg.* 2023;89:1622-8.
91. Sone K, Tanimoto S, Toyohara Y, Taguchi A, Miyamoto Y, Mori M, et al. Evolution of a surgical system using deep learning in minimally invasive surgery (Review). *Biomed Rep.* 2023;19:45.
92. Mao X, Jia P, Zhang L, Zhao P, Chen Y, Zhang M. An evaluation of the effects of human factors and ergonomics on health care and patient safety practices: a systematic review. *PLoS One.* 2015;10:e0129948.
93. Goggins RW, Spielholz P, Nothstein GL. Estimating the effectiveness of ergonomics interventions through case studies: implications for predictive cost-benefit analysis. *J Safety Res.* 2008;39:339-44.
94. Tetteh E, Wang T, Kim JY, Smith T, Norasi H, Van Straaten MG, et al. Optimizing ergonomics during open, laparoscopic, and robotic-assisted surgery: a review of surgical ergonomics literature and development of educational illustrations. *Am J Surg.* 2024;235:115551.